

Lot specific information

Lot number: 107-25351

Expiration date: 01-OCT-2027

Information

The product has been analyzed and is certified to have met the specifications.

Test	Results	
ATPlite 1step substrate: 760-25321		
Signal half-life	204	Minutes
Signal height	4.05 x 10 ⁶	CPS
Background	< 350	CPS
ATPlite 1step buffer: 750-25241		
pH	7.00	
Ratio	1.0	
Background	< 350	CPS

This Certificate of Analysis has been produced electronically and is valid without a signature.
 For questions regarding this Certificate of Analysis, please contact your customer service representative.