

Lot specific information

Lot number: 110-25431
 Expiration date: 01-APR-2028

Information

The product has been analyzed and is certified to have met the specifications.

Test		Results
britelite plus substrate: 660-25391		
Ratio luminescent light-output		1.00
Signal half-life		42 min
Ratio signal half-life		0.99
Background		< 75 cps
britelite plus buffer: 650-25391		
Ratio luminescent light-output		1.01
Signal half-life		42 min
Ratio signal half-life		0.99
Background		< 75 cps

This Certificate of Analysis has been produced electronically and is valid without a signature.
 For questions regarding this Certificate of Analysis, please contact your customer service representative.